




## Ocular Landers Wide Field Corneal Window

	Product Code	Image Mag	Contact Diameter	Static FOV		
 	OLTCW	2.3x	6-8.2mm	28°	<i>Designed with: Maurice B. Landers, III, M.D. Chapel Hill, NC</i>	

### Design

- § The Landers Wide Field Corneal Window is designed with a convex anterior surface to facilitate viewing of the peripheral retina & the posterior pole during vitrectomy corneal transplant surgery.
- § The design of the OLTCW allows for sealing trephine sizes 6-8.2mm in addition to irregular shaped corneal openings.
- § This lens has six suture holes around its peripheral edge. Lens design is suitable for phakic, aphakic and pseudophakic eye.
- § Once the sutures are in place, the lens/cornea interface acts as a seal and provides a closed system for a vitrectomy procedure.
- § May be used in conjunction with our handled vitrectomy lenses (OBVI, OFVI, OMVI, OPGVI, OPVI-3) or the Machemer Flat Vitrectomy Lens (OLV-5).

### Technique

- § A 7.0mm or 8.0mm corneal trephine is used to remove corneal button. The 7.0mm is most commonly used.
- § The lens is sutured to the cornea using 8.0 nylon suture.
- § A Flieringa ring should also be sutured to eyes with flexible scleras.
- § The large aperture provides a wide angle view of the posterior retina and permits clear visualization of the peripheral retina when combined with scleral depression.
- § In some cases, the fundus view can be improved by placing a Machemer Flat Vitrectomy Lens on the lens with a drop of visco-elastic material between the lenses.
- § At the conclusion of the pars plana vitrectomy, the sutures holding the lens are removed. The donor button is sutured in place.

### Cleaning & Disinfection

See Cleaning Method 1

